

# Calibration Certificate

Date: 18-April-2023

Certificate #: 70341

For:  
AMECO HOUSTON  
5200 GALVESTON RD

HOUSTON TX 77017

## Equipment Information

I.D.:	1075772	Serial Number:	0619110612
Manufacturer:	WRIGHT TOOLS	Model Number:	N/A
Unit Type:	TORQUE WRENCH (10-100 FT·LBF)	Cal Date:	4/18/2023
Temp./RH:	68 F / 50 % RH	Cal. Due Date:	4/18/2024
Cal. Interval:	12 MONTHS	Received condition:	OK FOR CAL
Calibration result:	PASS		

## Calibration Notes

IN TOLERANCE

## Test Points

Seq.	Description	Standard	UOM	Tolerance -	Tolerance +	As Found	As Left
1	20	20	FT·LBF	19	21	20	20
2	60	60	FT·LBF	58	62	60	60
3	100	100	FT·LBF	96	104	101	101

## Standards Used To Calibrate Equipment


I.D.	Manufacturer	Model	Description	Last Cal.	Cal. Due Date	Certificate Number
GCC-157	WILLIAMS	7065	TORQUE TESTER (500 FTLB)	5/26/2022	5/26/2023	57722

## Procedures Used In This Event:

Procedure Name	Revision Level	Revision Date
SCP-53		

Calibration Performed By  
BEN JOHNSON  
CALIBRATION TECHNICIAN

Approved By  
G. Lawal  
CEO



This unit met the measured values when it left our facility. No guarantees are made as to its reading in the field. Handling and usage can and may affect the accuracy of this unit.

These results relate only to the item calibrated and not to any other equipment.

This information may not be reproduced, except in full, without prior written consent.

Calibrations are performed using standards traceable to NIST. Our calibration system complies with ISO 17025.

A copy of this cert can be downloaded from <https://GulfCoastCalibration.IndySoft.com>. Contact your Sales Rep to get a name and password.