

Calibration Certificate

Date: 18-April-2023

Certificate #: 70339

For:
AMECO HOUSTON
5200 GALVESTON RD

HOUSTON TX 77017

Equipment Information

I.D.:	1075198	Serial Number:	1075198
Manufacturer:	WRIGHT TOOLS	Model Number:	N/A
Unit Type:	TORQUE WRENCH (30-250 FT·LBF)	Cal Date:	4/18/2023
Temp./RH:	68 F / 50 % RH	Cal. Due Date:	4/18/2024
Cal. Interval:	12 MONTHS	Received condition:	OK FOR CAL
Calibration result:	PASS		

Calibration Notes

IN TOLERANCE

Test Points

Seq.	Description	Standard	UOM	Tolerance -	Tolerance +	As Found	As Left
1	50						
2	150	50	FT·LBF	48	52	50	50
3	250	150	FT·LBF	144	156	152	152
		250	FT·LBF	240	260	255	255

Standards Used To Calibrate Equipment


I.D.	Manufacturer	Model	Description	Last Cal.	Cal. Due Date	Certificate Number
GCC-157	WILLIAMS	7065	TORQUE TESTER (500 FTLB)	5/26/2022	5/26/2023	57722

Procedures Used In This Event:

Procedure Name	Revision Level	Revision Date
SCP-53		

Calibration Performed By
BEN JOHNSON
CALIBRATION TECHNICIAN

Approved By
G. Lawal
CEO



This unit met the measured values when it left our facility. No guarantees are made as to its reading in the field. Handling and usage can and may affect the accuracy of this unit.
These results relate only to the item calibrated and not to any other equipment.
This information may not be reproduced, except in full, without prior written consent.
Calibrations are performed using standards traceable to NIST. Our calibration system complies with ISO 17025.
A copy of this cert can be downloaded from <https://GulfCoastCalibration.IndySoft.com>. Contact your Sales Rep to get a name and password.